



March 2021

Dear Valued BD Customer,

BD is happy to inform you that testing has been completed to support a shelf-life (expiration date) of up to 12 months for **BD Veritor™ System for Rapid Detection of SARS-CoV-2 kit, material number: 256089**. This letter is to notify you the BD Veritor™ System for Rapid Detection of SARS-CoV-2 kit 256089 in your possession may now have a longer than labeled product expiry date.

A listing of **BD Veritor™ System for Rapid Detection of SARS-CoV-2 kit, material number: 256089**, which qualify for this expiry extension, can be found in Attachment 1. These lots were originally dated with an expiry of 6 months, and data supports a 6-month extension (12 months total). The attachment lists the lot number, currently labeled kit expiry and new kit expiry date.

Please be aware that the product Unique Device Identifier (UDI) barcode on the kit box will display the original expiry date until you receive kits with the updated labeling.

We are pleased to offer our customers this additional dating in support of their COVID-19 testing efforts.

BD is committed to advancing the world of health. Our primary objectives are patient and user safety, and providing you with quality products.

For more information or any questions, please do not hesitate to contact your local BD representative.

Kind Regards,

A handwritten signature in black ink that reads "Carol Nieto".

Carol Nieto
Senior Director, Quality Management
BD Integrated Diagnostic Solutions



Attachment 1 - BD Veritor™ System for Rapid Detection of SARS-CoV-2 kit,
 material number: 256089 lots with extended expiry

bd.com

Lot Number	Original Expiry	New Expiry
0294803	3/25/2021	09/25/2021
0295733	3/26/2021	09/26/2021
0331698	4/1/2021	10/01/2021
0332652	4/2/2021	10/02/2021
0311781	4/6/2021	10/06/2021
0303674	4/7/2021	10/07/2021
0314005	4/8/2021	10/08/2021
0309513	4/9/2021	10/09/2021
0331906	4/12/2021	10/12/2021
0333772	4/13/2021	10/13/2021
0342239	4/16/2021	10/16/2021
0335228	4/19/2021	10/19/2021
1013902	5/17/2021	11/17/2021
1010086	5/18/2021	11/18/2021
0340740	5/20/2021	11/20/2021
0343900	5/20/2021	11/20/2021
0346215	5/22/2021	11/22/2021
0349155	5/26/2021	11/26/2021
0357565	5/31/2021	11/31/2021
0358634	6/2/2021	12/02/2021
0356935	6/4/2021	12/04/2021
0364277	6/7/2021	12/07/2021
0365333	6/9/2021	12/09/2021
0361837	6/10/2021	12/10/2021
1002327	6/11/2021	12/11/2021
1003502	6/14/2021	12/14/2021
1007570	6/16/2021	12/16/2021
1013906	6/22/2021	12/22/2021
1017995	6/22/2021	12/22/2021
1014710	6/24/2021	12/24/2021
1022146	6/24/2021	12/24/2021
1016642	6/25/2021	12/25/2021
1020841	6/25/2021	12/25/2021
1019883	6/29/2021	12/29/2021
1018218	6/30/2021	12/30/2021
1026113	6/30/2021	12/30/2021

BD, the BD Logo and Veritor are trademarks of Becton, Dickinson and Company or its affiliates. © 2021 BD. All rights reserved.

Unless otherwise indicated, all products are CE-marked in compliance with the European In Vitro Diagnostic Medical Devices Directive 98/79/EC.



1021217	7/1/2021	01/02/2022
1022997	7/2/2021	01/02/2022
1024065	7/4/2021	01/04/2022
1029376	7/6/2021	01/06/2022
1032077	7/6/2021	01/06/2022
1035279	7/6/2021	01/06/2022
1027050	7/7/2021	01/07/2022
1033798	7/7/2021	01/07/2022
1035122	7/7/2021	01/07/2022
1028825	7/9/2021	01/09/2022
1031666	7/9/2021	01/09/2022
1036924	7/10/2021	01/10/2022
1046484	7/10/2021	01/10/2022
1038934	7/12/2021	01/12/2022
1038689	7/13/2021	01/13/2022
1041056	7/15/2021	01/15/2022
1042874	7/15/2021	01/15/2022
1040092	7/16/2021	01/16/2022
1045248	7/16/2021	01/16/2022
1044778	7/19/2021	01/19/2022
1047008	7/19/2021	01/19/2022
1050036	7/19/2021	01/19/2022
1050033	7/19/2021	01/19/2022
1047060	7/20/2021	01/20/2022
1050617	7/20/2021	01/20/2022
1047054	7/21/2021	01/21/2022
1047064	7/21/2021	01/21/2022
1050591	7/21/2021	01/21/2022

